

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 01/16/2023

OPERATOR NAME:
 DAMIAN E DUFFY
 Permit #: 49487
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:37
INTERNAL STANDARD	VERIFIED	03:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:38
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:40
BLANK TEST	0.00	03:41

