

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 01/08/2023

OPERATOR NAME:
DAVID M COSTELLO
Permit #: 40891
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:01
INTERNAL STANDARD	VERIFIED	07:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05

