

SOLUTION CHANGE

 WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 01/02/2023

OPERATOR NAME:
 DAVID F BUXTON
 Permit #: 33398
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:38
INTERNAL STANDARD	VERIFIED	09:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42

