

SUPERVISOR TEST

-----  
 WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806  
 -----

Date: 01/21/2023

OPERATOR NAME:  
 BRYAN J GRECO  
 Permit #: 44077  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:16
INTERNAL STANDARD	VERIFIED	16:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:17
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:18
BLANK TEST	0.00	16:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:19
BLANK TEST	0.00	16:20

