

SUPERVISOR TEST

WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806

Date: 01/15/2023

OPERATOR NAME:  
 ALEXANDER T GILMORE  
 Permit #: 47649  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:06
INTERNAL STANDARD	VERIFIED	06:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:07
BLANK TEST	0.00	06:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:09
BLANK TEST	0.00	06:10

