

SUPERVISOR TEST

WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806

Date: 01/07/2023

OPERATOR NAME:  
 DAVID F BUXTON  
 Permit #: 33398  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:03
BLANK TEST	0.00	14:04

