

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/26/2022

OPERATOR NAME:
TIMOTHY M GILLIS
Permit #: 51160
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:59
INTERNAL STANDARD	VERIFIED	22:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:59
BLANK TEST	0.00	23:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:01
BLANK TEST	0.00	23:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:02
BLANK TEST	0.00	23:03

