

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/02/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:00
INTERNAL STANDARD	VERIFIED	16:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:01
BLANK TEST	0.00	16:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:03
BLANK TEST	0.00	16:04

12/02/2022


