

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/12/2022

OPERATOR NAME:
BRIANNNA M JONES
Permit #: 51164
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:11
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:13

