

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 12/12/2022

OPERATOR NAME:  
MICHAEL O WEBER  
Permit #: 46067  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:01
INTERNAL STANDARD	VERIFIED	08:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:01
BLANK TEST	0.00	08:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:02
BLANK TEST	0.00	08:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:03
BLANK TEST	0.00	08:04

