

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/26/2022

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:58
INTERNAL STANDARD	VERIFIED	16:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:58
BLANK TEST	0.00	16:59

RFI Detected

