

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/26/2022

OPERATOR NAME:
MATHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:02
INTERNAL STANDARD	VERIFIED	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:02
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05

