

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/16/2023

OPERATOR NAME:
ERIC C WRIGHT
Permit #: 49047
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:38
INTERNAL STANDARD	VERIFIED	06:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:39
BLANK TEST	0.00	06:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:40
BLANK TEST	0.00	06:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:41
BLANK TEST	0.00	06:42

