

SUPERVISOR TEST

-----  
WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406  
-----

Date: 01/09/2023

OPERATOR NAME:  
JEFFREY E PETERMAN  
Permit #: 49748  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:02
INTERNAL STANDARD	VERIFIED	07:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06

